



National Accreditation Board for
Testing and Calibration Laboratories

CERTIFICATE OF ACCREDITATION

**PLA ELECTRO APPLIANCES PVT. LTD. (CALIBRATION
LAB)**

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2017

**"General Requirements for the Competence of Testing &
Calibration Laboratories"**

for its facilities at

PLA HOUSE, THAKOR INDUSTRIAL ESTATE, KURLA - KIROL ROAD, MUMBAI, MAHARASHTRA, INDIA

in the field of

CALIBRATION

Certificate Number: CC-3048

Issue Date: 09/11/2022

Valid Until:

08/11/2024

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Name of Legal Identity : PLA ELECTRO APPLIANCES PRIVATE LIMITED

Signed for and on behalf of NABL



N. Venkateswaran
Chief Executive Officer



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : PLA ELECTRO APPLIANCES PVT. LTD. (CALIBRATION LAB), PLA HOUSE, THAKOR INDUSTRIAL ESTATE, KURLA - KIROL ROAD, MUMBAI, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-3048 **Page No** 1 of 1

Validity 09/11/2022 to 08/11/2024 **Last Amended on** -

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	RADIOLOGICAL - RADIOLOGICAL MEASUREMENT S	Pocket Dosimeters	AERB Standardization Report & SSD, AERB/BARC methods	1 µSv to 10 Sv	7.5%
2	RADIOLOGICAL - RADIOLOGICAL MEASUREMENT S	Radiation Survey Instruments (Survey Meters, Area Monitors and Contamination Monitors) Pocket Dosimeters	AERB Standardization Report & SSD, AERB/BARC methods	0.5 mR/h to 5 R/h	8.9%
3	RADIOLOGICAL - RADIOLOGICAL MEASUREMENT S	Radiation Survey Instruments (Survey Meters, Area Monitors and Contamination Monitors) Pocket Dosimeters	AERB Standardization Report & SSD, AERB/BARC methods	10 mR/h to 500 R/h	7.5%

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.